


## Instructions for Use

# Traction Boot Liners


40531H

Refer to your facility's patient positioning protocols prior to use

1. Establish proper orientation of the boot liner. The sewn edge will align with the sole of the patient's foot.
2. Open the Velcro closure straps and slide the open boot liner over the patient's foot in the proper orientation.
3. Wrap the sides of the boot liner around the foot and secure the Velcro closure. The fit should be snug without impeding circulation.
4. Wrap the lower calf portion of the leg and secure the remaining Velcro closure. Again, be sure to not over-tighten.
5. The boot liner is now secured and the foot is ready for placement into the traction boot. Repeat steps 1-4 for the remaining foot.




- 1 Sewn edge aligns to sole of foot.




2



3



4




5

LITIFUPP032 R1 5/27/2021



702 Prominence Drive  
Pittsburgh, PA 15068 U.S.A.

Tel: 724.337.5500  
Fax: 724.337.0555



xodusmedical.com  
info@xodusmedical.com

FDA REGISTERED  
ISO 13485 CERTIFIED

  
xodusmedical.com/eifu

EMERGO EUROPE  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands